

Sample Statistical Summary

MAPEP-11-GrF25

Analyte	T(1)	A(2)	Grand(3) Mean	Std Dev	Ref Value	Ref Unc	Acceptance Range
Radiological							Units: (Bq/sample)
Gross alpha	57	47					False Positive Test
Gross beta	59	33					False Positive Test

Note: (1) T = Total number of laboratories reporting analyte.
(2) A = Number of laboratories with 'Acceptable' performance.
(3) Mean excludes values derived as total metals and values indicated as "Not Acceptable".

Gross Alpha Flags:

A = Result acceptable, Bias 2, i.e., the range encompassing the result, plus or minus the total uncertainty at two standard deviations, does not include zero).

N = Result not acceptable, Bias > +/- 70% or the reported result is not statistically positive at two standard deviations (Result/Uncertainty)

Gross Beta Flags:

A = Result acceptable, Bias 2, i.e., the range encompassing the result, plus or minus the total uncertainty at two standard deviations, does not include zero).

N = Result not acceptable, Bias > +/- 50% or the reported result is not statistically positive at two standard deviations (Result/Uncertainty)